

Central Lancashire Online Knowledge (CLOK)

Title	Determination of meningioma brain tissue grades using Raman hyperspectral imaging
Type	Article
URL	https://clock.uclan.ac.uk/id/eprint/30273/
DOI	https://doi.org/10.1093/neuonc/noz167.022
Date	2019
Citation	Medeiros-De-morais, Camilo De lelis orcid iconORCID: 0000-0003-2573-787X, Lilo, Taha, Ashton, Katherine, Davis, Charles, Dawson, Timothy, Gurusinghe, Nihal and Martin, Francis L (2019) Determination of meningioma brain tissue grades using Raman hyperspectral imaging. <i>Neuro-Oncology</i> , 21 (S4). pp. 5-6. ISSN 1522-8517
Creators	Medeiros-De-morais, Camilo De lelis, Lilo, Taha, Ashton, Katherine, Davis, Charles, Dawson, Timothy, Gurusinghe, Nihal and Martin, Francis L

It is advisable to refer to the publisher's version if you intend to cite from the work.
<https://doi.org/10.1093/neuonc/noz167.022>

For information about Research at UCLan please go to <http://www.uclan.ac.uk/research/>

All outputs in CLOK are protected by Intellectual Property Rights law, including Copyright law. Copyright, IPR and Moral Rights for the works on this site are retained by the individual authors and/or other copyright owners. Terms and conditions for use of this material are defined in the <http://clock.uclan.ac.uk/policies/>

Determination of meningioma brain tissue grades using Raman hyperspectral imaging

Camilo L. M. Morais^{1*}, Taha Lilo², Katherine M. Ashton², Charles Davis², Timothy P. Dawson², Nihal Gurusinghe², Francis L. Martin¹

¹School of Pharmacy and Biomedical Sciences, University of Central Lancashire, Preston PR1 2HE, UK

²Royal Preston Hospital, Lancashire Teaching Hospitals NHS Trust, Preston PR2 9HT, UK

[*cdlmedeiros-de-morai@uclan.ac.uk](mailto:cdlmedeiros-de-morai@uclan.ac.uk)

Raman spectroscopy is a powerful tool used to analyse biological materials, where spectral biomarkers such as proteins (1500-1700 cm^{-1}), carbohydrates (470-1200 cm^{-1}) and phosphate groups of DNA (980, 1080-1240 cm^{-1}) can be detected. A major advantage is that it is reagent-free and unaffected by water interference, which is ideal for biological applications. Raman hyperspectral imaging combines the chemical sensitivity of this spectrochemical technique with spatially distributed information. Herein, Raman microspectroscopy imaging ($50 \times 50 \mu\text{m}$ tissue area, $50\times$ magnification, 50% laser power, 0.1 ms exposure time, 780-1858 cm^{-1} spectral range) was used to investigate 79 brain tissue samples (sourced from the Brain Tumour North West) in order to differentiate meningioma Grade I ($n=55$) *versus* Grade II ($n=24$). Meningioma is the commonest type of brain tumour with the majority of them being benign tumours (Grade I) whilst a few are aggressive or malignant (Grade II). Grade II tumours have a poor prognosis by their nature, hence new sensitive diagnostic tools are essential. Using partial least squares discriminant analysis (PLS-DA) with 20 LVs (99% explained variance), we were able to differentiate Grade I *versus* Grade II meningioma with a classification accuracy of 99% in an internal validation set of 23 samples (16 Grade I, seven Grade II); this was with an accuracy of 85% for an external test set of similar size. These findings highlight the potential of Raman hyperspectral imaging for differentiation of meningioma tumours.