

Central Lancashire Online Knowledge (CLoK)

Title	The use of computer-aided design and manufacture for foot orthoses: a cross-sectional study of orthotic services in the UK
Type	Article
URL	https://clock.uclan.ac.uk/id/eprint/53622/
DOI	https://doi.org/10.1002/jfa2.70031
Date	2025
Citation	Barr, Laura, Richards, James and Chapman, Graham (2025) The use of computer-aided design and manufacture for foot orthoses: a cross-sectional study of orthotic services in the UK. Journal of Foot and Ankle Research, 18 (1). e70031.
Creators	Barr, Laura, Richards, James and Chapman, Graham

It is advisable to refer to the publisher's version if you intend to cite from the work.
<https://doi.org/10.1002/jfa2.70031>

For information about Research at UCLan please go to <http://www.uclan.ac.uk/research/>

All outputs in CLoK are protected by Intellectual Property Rights law, including Copyright law. Copyright, IPR and Moral Rights for the works on this site are retained by the individual authors and/or other copyright owners. Terms and conditions for use of this material are defined in the <http://clock.uclan.ac.uk/policies/>

Figure 5. The most common workflow for CAD/CAM insole production in UK Orthotic services:




Shape capture - 86.8% (79/91) of services most commonly used foam box impression casts when manufacturing CAD/CAM insoles

The patient's foot is pressed into a phenolic foam box to produce a negative cast of their foot



Scanning - 90.8% (79/87) of services reported that the casts were transported and scanned into the CAD/CAM system at another site

The foam-box cast is physically transported to a site with a digital scanner



Digital upload - 81.6% (71/87) of services scanned the negative cast into the CAD/CAM system

The negative foam-box cast is scanned directly into the digital scanner without being filled with plaster



Insole design – in 73.6% (67/91) of services a technician modelled the CAD/CAM insole

The digital foot model is modified in the CAD/CAM system in order to create the final insole design



Manufacture – mean 59.1% (SD 37.92) of all CAD/CAM insoles were produced by reduction milling

The physical insole is produced by a milling machine from the digital insole design which is milled into a block of material such as ethylene-vinyl acetate